

SELF-EXPANDING STENT GRAFTS AVAILABLE FOR SALE IN THE US

Company Name	Device Name	Indicated Use	Material Used	Introducer Size (F)	Endhole (inch)	Stent Diameter (mm)	Stent Length (mm)	Delivery System Length (cm)
Bard Peripheral Vascular	Fluency Plus	Tracheobronchial	Nitinol/ ePTFE	8–9	.035	6, 7, 8, 9, 10	40, 60, 80	80, 117
Boston Scientific Corporation	Wallgraft	Tracheobronchial	PET-covered Elgiloy	9, 10, 11, 12	.035	6, 7, 8, 9, 10, 12, 14	27, 28, 29, 30, 41, 43, 44, 68, 71, 72, 74, 96, 98, 100, 101, 104	90
Gore & Associates	Viabahn, Viabahn with Heparin Bioactive Surface	5, 6, 7, 8 indicated for SFA; 9, 10, 11, 13 cleared for tracheobronchial malignant neoplasms	Nitinol/ ePTFE	7–12	.035–.025	5, 6, 7, 8, 9, 10, 11, 13	25, 50, 100, 150	75, 110, 120
	Viabil	Malignant biliary strictures		10	.035	8, 10	40, 60, 80, 100	40, 195
	Viatorr	TIPS, tracheobronchial			.038	8, 10, 12	40, 50, 60, 70, 80	75
LeMaitre Vascular, Inc.	aSpire	Tracheobronchial malignant neoplasms	ePTFE covered nitinol	7	.018	6, 7, 8, 9	25, 50, 100, 150	50, 100